


FOUNDRY DIVISION


Date : 10-12-2013


Description of Material : THERMOCOUPLE TIPS BMD 13%
Annual Quantity : 1500 Nos.
Approx. Annual Value : Rs. 90,000/-


Pre-qualification criteria for supply of Thermocouple Tips BMD 13%

1. Supplier should be manufacturer / agent of Thermocouple Tips BMD Pt v/s Pt 13% Rh.
2. Thermocouple Tips BMD 13% is use in temperature measurement of molten metal (500 - 1810 °C). Thermocouple Tips design for multiple dips in slag free molten metal. Tube length of tips 70 mm and no sputtering and spilling occur when dips in molten metal.
3. Party should have Monthly capacity of Minimum 1000 Nos. for supply of Thermocouple Tips BMD 13%.
4. The party should have the experience of minimum three years of supplying Thermocouple Tips BMD 13%. Work order / P.O. to be submitted for proof.
5. List of customers along with contact details to be furnished.
6. Following must be acceptable to interested parties :-
 - BHEL reserves the right to inspect the material at site before dispatch. The supplier shall intimate in advance enclosing a copy of test certificate of the material offered for inspection.
 - If the material received is not found suitable to our requirements, it shall be replaced free of cost by supplier.
 - Final tests shall be carried out at BHEL and material will be accepted or rejected based on this result.

A.K. Singh (Sr. Mgr., FYM) 

 Abhijit Sinha
अभिजीत सिंहा | Abhijit Sinha
अभिजीत सिंहा एवं विभागाध्यक्ष | A.G.M. & H.O.D.
अभिजीत सिंहा | Foundry Division
बी.एच.एल., भोपाल | B.H.E.L., BHOPAL

 Anil Kumar Singh
अनिल कुमार सिंहा | Anil Kumar Singh
अनिल कुमार सिंहा | Manager (Prod.)
अनिल कुमार सिंहा | Foundry Division
बी.एच.एल., भोपाल | B.H.E.L., BHOPAL

 Ashwani Singh
Asst. Engineer (FYM)